

## **Kohler-ALBACORE LC4x4 Generalized Response and Calibration Factor**

These calculations are for the generalized case and assume the signal is in the sensor frequency range giving a flat response. Frequency response ranges are indicated.

### SENSOR RESPONSE INFO:

For the Differential Pressure Gauge (DPG) sensitivity:

Using measurements from strain gauge full scale outputs, the average is about 57 mv/7e3 PA. (FSO is 75e-3V/7e3PA). According to Cox *et al* there is a mechanical attenuation factor due to the compressibility of the oil and the compliance of the chamber. They estimate this as 0.86. Willoughby *et al* say they measured the response and deduced this factor as 0.924. Using a value of 0.9 is suggested until a better value can be determined. Thus the STAGE 1 gain should be  $5.7e-2 * 0.9 / 7e3$  or:

**S(DPG) = 7.3  $\mu$ V/Pa      -or-      (7.3e-6 V/Pa)**

*flat response: 0.002 Hz (500 sec) to ~30+ Hz*

Frequency response information:

| Parameter               | Nominal Value | Units |
|-------------------------|---------------|-------|
| Zeros (1)               | 0             | Rad/s |
| Poles (1)*              | -0.012568     | Rad/s |
| Normalization           | 1             |       |
| Normalization Frequency | 0.3           | Hz    |

\* Single Pole @ (1/-79.57).

For the High-Tech Hydrophone sensitivity:

For the custom High-Tech Hydrophone (HTI-90-U) the manufacturer calibration files give a sensitivity of -182.7 dB re 1V/ $\mu$ Pa. This hydrophone loses ~2 dB in sensitivity per ~6000m in depth (10,000 psi) so for typical ocean depth around 3km we correct ~1 dB and use -183.7 dB re 1V/ $\mu$ Pa. Using amplitude spectra throughout (e.g.  $X[db] = 20 * \log_{10}[X/Xref]$ ), this gives  $S(hyd-HTI) = 10^{**(-183.7/20)} * 1V/\mu Pa = .653 mV/Pa$  (@ 3000m water depth). Thus:

**S(hyd-HTI) = 0.653 mV/Pa    -or-    (6.53e-4 V/Pa)**

*flat response: 0.05 Hz to 7.5 kHz (@ 3000m depth)*

Frequency response information:

From Brian Spychalski at High Tech Inc. (personal communication: June 03, 2014):

- 1: HPF at input of preamp created by ceramic element 12.8nF (nom.) and 300 meg ohm resistor (0.04Hz)
- 2: There is another HPF at opamp set at 0.02Hz.
- 3: LPF at opamp is set at 7.5KHz.

| Parameter               | Nominal Value                      | Units |
|-------------------------|------------------------------------|-------|
| Zeros (2)               | 0 0                                | Rad/s |
| Poles (3)               | -24.127431<br>-0.1256637<br>-47124 | Rad/s |
| Normalization           | 47124                              |       |
| Normalization Frequency | 500                                | Hz    |

For the Trillium-240-OBS seismometer sensitivity:

the manufacturer quotes 1196.5 V\*s/m over +/-20V for a full differential signal. SIO-4x4-LP uses only a single-sided input to the A/D, effectively halving the sensitivity, thus:

**S(T240-ss) = 598.25 V/m/s**

*flat response: 0.004167 Hz (240 sec) to 35 Hz*

Trillium 240 OBS Seismometer Frequency response information:  
(From Trillium 240 OBS User Guide - page 10)

**Table 3-2** Poles and zeroes

|              | <b>Parameter</b>            | <b>Nominal values</b>   | <b>Units</b> |
|--------------|-----------------------------|-------------------------|--------------|
| $z_n$        | Zeroes                      | 0                       | rad/s        |
|              |                             | 0                       |              |
|              |                             | -108                    |              |
|              |                             | -161                    |              |
| $p_n$        | Poles                       | -0.01815 ±0.01799i      | rad/s        |
|              |                             | -173                    |              |
|              |                             | -196 ±231i              |              |
|              |                             | -732 ±1415i             |              |
| $k$          | Normalization factor        | 2.316 x 10 <sup>9</sup> |              |
| $S_{sensor}$ | Passband sensitivity at 1Hz | 1196.5                  | V*s/m        |
| $f_0$        | Normalization frequency     | 1                       | Hz           |

\*\*NOTE: Sensor sensitivity listed in Table 3-2 is for full differential response; SIO uses single sided input (halving this number to 598.25 V\*s/m).

For the Trillium-40 seismometer sensitivity:  
 the manufacturer quotes 1553 V\*s/m over +/-8V for a full differential signal. SIO-LC4x4 uses only a single-sided input to the A/D, effectively halving the sensitivity, thus:

**S(T40-ss) = 776.5 V/m/s** *flat response: 0.025 Hz (40 sec) to 50 Hz*

Trillium 40 Seismometer Frequency response information:  
 (From Trillium 40 User Guide - page 40)

**Table 9-1** Ground motion response nominal parameters

| Symbol | Parameter                          | Nominal Values         | Units                |
|--------|------------------------------------|------------------------|----------------------|
| $z_n$  | Zeros                              | 0                      | rad/s                |
|        |                                    | 0                      |                      |
|        |                                    | -68.8                  |                      |
|        |                                    | -323                   |                      |
|        |                                    | -2530                  |                      |
| $p_n$  | Poles                              | -0.1103 ± 0.1110i      | rad/s                |
|        |                                    | -86.3                  |                      |
|        |                                    | -241 ± 178i            |                      |
|        |                                    | -535 ± 719i            |                      |
| $k$    | Normalization factor               | 1.104x 10 <sup>5</sup> | (rad/s) <sup>2</sup> |
| $f_0$  | Normalization frequency            | 1                      | Hz                   |
| $S$    | Ground motion sensitivity at $f_0$ | 1553                   | V s/m                |

\*\*NOTE: The passband sensitivity listed in Table 4-2 is for full differential response, SIO uses single sided input (halving this number to 776.5 V\*s/m).

For the L28LB seismometer sensitivity:

Transduction constant --> 1.57 \* sqrt(R-coil) V/m/s with R-coil = 630 ohm nominally this gives 39.53 V/m/s. SIO uses 70% coil current damping, (R-shunt-ss = (7860+51) ohm single-sided, divide by 2 to effective R-shunt damping for differential signal), thus R-shunt-diff = 3956 ohm, which gives:

**S(L28) = 34.10 V/m/s** *flat response: ~4.5 Hz and above*

Frequency response information (for a damped oscillator):

Two zeros at 0, two poles at  $\omega_0 \left( \delta \pm i\sqrt{1-\delta^2} \right)$  where  $\delta = 0.701$  (damping value).

| Parameter               | Nominal Value       | Units |
|-------------------------|---------------------|-------|
| Zeros (2)               | 0 0                 | Rad/s |
| Poles (2)               | 19.820 +/- i*20.164 | Rad/s |
| Normalization           | -1                  |       |
| Normalization Frequency | 4.5                 | Hz    |

**ELECTRONICS RESPONSE INFO:**

The sensitivity of the A/D is as follows:

With reference filter voltage of V-filt = 100 ohm the voltage range is +/- 2.47 V, max counts over this range are -Vref = -6,100,300 to +Vref = 6,102,081.

This gives  $S(a/d) = 4.94 / 12,202,381 = 0.405 \times 10^{-6} \text{ V/count} = 0.405 \text{ microV/count}$ , or:

**$S(a/d) = 0.405 \text{ } \mu\text{V/count} \quad \text{-or-} \quad (4.05\text{e-}7 \text{ V/count})$**

*Note: A/D reaches full 24-bit range (i.e. -8388608 to 8388607) @ overvoltage of +/- 3.3 V. However, the response in this overvoltage range is roughly nonlinear.*

*Note2: If V-filt = 10 ohm the voltage range is +/- 2.50 V →  $S(a/d) = 0.410 \text{ microV/count}$ .*

**PRE-AMP GAIN INFO:**

Pre-amp gain settings for sensor/channel on all LC4x4 OBS deployments are:

- gain(DPG) = 64**
- gain(hyd-HTI) = 1**
- gain(T240-ss) = 0.102**
- gain(T40-ss) = 0.200**
- gain(L28) = 64**

*Note: To keep the Trillium on scale at the A/D input (max +/- 2.47 V), signal from the Trillium seismometers output are attenuated using an analog voltage divider on the pre-amp board:*

*$V\text{-T240-div} = R\text{-T240-gnd-eff} / (R\text{-T240-sig} + R\text{-T240-gnd-eff}) = 795 / (6980 + 795) = 0.102$*

*$V\text{-T40-div} = R\text{-T40-gnd-eff} / (R\text{-T40-sig} + R\text{-T40-gnd-eff}) = 1746 / (6980 + 1746) = 0.200$*

**TOTAL SYSTEM RESPONSE INFO:**

Total system response then becomes:  $S(\text{total}) = S(a/d) / [S(\text{sensor}) * \text{gain}]$

**LC4x4 Generalized Total System Response:**

**LC4x4-LP units:**

**DPG pressure response = 0.867 mPa/count (500 sec to ~30+ Hz)**  
**= 8.67e-4 Pa/count**

**Trillium-240-OBS Velocity response = 6.637 (nm/s)/count (240 sec to 35 Hz)**  
**= 6.637e-9 (m/s)/count**

**Trillium-40-OBS Velocity response = 2.608 (nm/s)/count (40 sec to 50 Hz)**  
**= 2.608e-9 (m/s)/count**

**LC4x4-SP units:**

**Hydro-HTI pressure response = 0.620 mPa/count (~0.05 Hz to 7.5 kHz)**  
**= 6.20e-4 Pa/count**

**L28 Velocity response = 0.186 (nm/s)/count (~4.5 Hz and above)**  
**= 1.86e-10 (m/s)/count**